


- Binding Buffer (600µl)
Columns 1 & 7
- Wash #2 Buffer (600µl)
Columns 4 & 10, 5 & 11
- Wash #1 Buffer (600µl)
Columns 2 & 8, 3 & 9
- Elution Buffer (100µl)
Columns 6 & 12



- ▲ 400µl sample added to Binding Buffer
- ▼▲ 100µl extracted and purified RNA elution
- ▲ 400µl sample added to Binding Buffer
- ▼ 100µl extracted and purified RNA elution


PRKit-A miQron protocol parameters

Step Name	Column	Volume (µl)	Time (sec)	Mixing Speed (1-10)	Dry Time (sec)	Magnet Capture Time (sec)
Binding	1 & 7	600	300	7	0	150
Wash #1	2 & 8	600	60	7	0	90
Wash #1	3 & 9	600	60	7	0	90
Wash #2	4 & 10	600	60	7	0	90
Wash #2	5 & 11	600	60	7	300	90
Elution	6 & 12	100	60	10	0	150
Discard Comb	2 & 8	600	0	5	0	0


1 Add up to 50mg of fresh plant leaves sample to the lysis bead tube provided.




2 Add 600µl Lysis Buffer, and 60µl PPB. Mix for 10 mins using TissueLyser at max speed or vortex for 10 mins; then centrifuge at 20,000g for 2 mins.




3 To the RNA Extraction Kit Plate **A** transfer up to 400µl of supernatant to Binding Buffer #1 (columns 1 & 7). You can add up to 16 samples.




4 Place plate into the miQron, taking care that the label is facing outward.




5 Insert two combs.



6 Select PRKit - Part A and press **▶**



When program is complete, remove plate from miQron and discard combs.




▲ 100µl sample added to Binding Buffer
▼ 50µl extracted and purified RNA elution
▲ 100µl sample added to Binding Buffer
▼ 50µl extracted and purified RNA elution

- Binding Buffer (400µl)
Columns 1 & 7
- Functionalized Beads (200µl)
Columns 2 & 8
- Wash #3 Buffer (600µl)
Columns 3 & 9
- Elution Buffer (50µl)
Columns 5 & 11


PRKit-B miQron protocol parameters

Step Name	Column	Volume (µl)	Time (sec)	Mixing Speed (1-10)	Dry Time (sec)	Magnet Capture Time (sec)
Bead Transfer	2 & 8	200	n/a	n/a	0	100
Binding	1 & 7	400	12	7	0	110
Wash #3	3 & 9	600	12	6	0	100
Elution	5 & 11	50	12	7	0	200
Discard Comb	3 & 9	600	0	5	0	0


7 Add 100µl of DNase Reaction Buffer to DNase pellet and then add 100µl of glycerol. Mix by gently inverting the tube. Reconstituted pellet must be stored at -20 °C.




8 For each RNA sample, prepare DNase Buffer by adding 10µl of DNase prepared in the previous step to 40µl of DNase Reaction Buffer in a microfuge tube. Mix DNase with buffer by gently inverting the tube a few times.




9 To the RNA Extraction Kit Plate A add 50µl of DNase Buffer to Elution Buffer (columns 6 & 12). Dispense directly into the elution buffer not on the walls. If there are any droplets of elution buffer left on the walls, slide them to the well bottom by gently shaking or tapping the plate on a solid surface.




10 Gently shake Plate A for 10 seconds by hand and incubate at room temp for 20 mins. From the RNA Extraction Kit Plate A transfer 100µL of the Elution Buffer (columns 6 & 12) to each well (columns 1 & 7) of the RNA Extraction Kit Plate B.




11 Place Plate B into the miQron, taking care that the label is facing outward.



12 Insert two combs.



13 Select PRKit - Part B and press ▶



When program is complete, remove plate from miQron and discard combs.

Columns 5 and 11 contain the purified RNA elution.

